



ADV Molecular Q Panel 01

ADV MQP01-A
GTIN: 5060605500133



Qnostics, Block 1 Todd Campus, West of Scotland Science Park
Glasgow, G20 0XA, Scotland, UK
www.qnostics.com



WMDE B.V. Bergerweg 18, 6085 AT Horn, The Netherlands



The following instructions must be read before using this product



Intended Use

Qnostics' Adenovirus (ADV) Molecular Q Panels are intended to help laboratories monitor their molecular assays on a run to run basis within customer derived limits.

The Panels can be used to support laboratory staff training and in the assessment and development of ADV molecular diagnostic assays from extraction phase through amplification to detection.

Principles of the Panel

The Panels are manufactured to ISO standard 13485:2016 compliant systems. The samples were produced by making quantitative dilutions of whole ADV type 1 into transport medium (see Table 1).

Samples are representative of clinical human specimens which are traceable to an internal reference preparation in line with the requirements of ISO 17511:2003.

The Panels are suitable for use with the majority of commercial and in-house molecular methods. They can also be used to support the training and monitoring of new operators in line with laboratories quality management requirements. The target concentrations of the Panel have been assigned to cover the dynamic range of most molecular assays and are consistent within each lot and across batches.

Product Description and Performance Characteristics

The Panel consists of 4 x 1ml samples, 3 positive and 1 negative. Panels are manufactured as a log (1:10) dilution series (see Table 1). The Panels are provided as liquid-frozen in a 'single-use' tube format and must be extracted immediately after thawing.

Table 1: Panel Components and Characteristics

Sample Code	Target Concentration Log ₁₀ IU/ml 1 st IS*	Target Concentration IU/ml	Number of vials
ADV MQP01-H	5.4	250,000	1
ADV MQP01-M	4.4	25,000	1
ADV MQP01-L	3.4	2,500	1
ADV MQP01-N	Negative	Negative	1

*ADV NIBSC 16/324

IMPORTANT NOTE: The Panel members have **no assigned value**. The values provided in Table 1 are specific to the Qnostics' reference assay used for the qualification of the Panel. The actual panel member quantification values may vary from those reported and are dependent on the evaluation procedure, the nucleic acid extraction and molecular assay used (see Limitations). It is the responsibility of the end user to establish their own target results for each of the Panel members using their laboratory's molecular procedures for their specific molecular diagnostic assay and appropriate statistical control.

Warnings and Precautions

The Panel contain whole ADV and must only be handled by trained laboratory personnel and in accordance with Good Laboratory Practices, which must include the use of personal protective equipment (PPE). All residual materials must be treated as potentially hazardous and disposed of accordingly. This must be carried out according to the established procedures of the laboratory and in accordance with national and international regulations.

Do not pipette by mouth. Do not eat, drink or smoke when handling the samples or within laboratory spaces. Observe the expiration date for the Panels.

Hazard and Precautionary Statements: H303, H333, P202, P270, P280, P314

Additional Equipment Required but not Provided

The following equipment is not included:

- Personal Protective Equipment (PPE) - e.g. lab coat and gloves
- Biological safety cabinet
- Nucleic acid extraction kit used in accordance with the manufacturers' instructions
- Molecular Amplification assay specific for ADV and, where appropriate, used in accordance with the manufacturers' instructions
- Bench vortex
- Micro-centrifuge (12-14,000 RPM)
- Calibrated pipettes and sterile barrier filter tips

Procedure

- The Panel must be thawed at room temperature.
- Vortex briefly and spin down at 12,000 RPM for 30 seconds before opening the sample tube.
- The samples must then be treated in the same manner to that required by the laboratory for routine specimens, in the normal ADV molecular procedure being assessed.

IMPORTANT NOTICE: Each panel member is intended for 'single use' ONLY. After thawing and testing any surplus material must be disposed of according to laboratory procedures.

For technical queries please contact info@qnostics.com

Storage

The Panel must be stored within the recommended temperature range of -20/-80°C.

The re-freezing and repeated thawing or off label storage of the Panel is not recommended and may lead to variability in the results obtained.

Limitations

The Panel **must not** be used as a substitute for assay process controls and / or calibrators (Standards) provided by the manufacturer of the molecular assay.

This product is not an absolute reference material. The laboratory needs to establish its own target results using the Panel for their particular molecular assay system.

References

World Health Organisation (WHO). Laboratory Biosafety Manual, 3rd ed. 2004 ISBN 92 4 154650 6 (LC/NLM classification: QY 25).

Centers for Disease Control (CDC). Recommendations for the prevention of HIV transmission in healthcare settings. MMWR 1987; 36, Supplement no. 2S.

Centers for Disease Control (CDC). Update: Universal guidelines for the prevention of transmission of human immunodeficiency virus, hepatitis B virus and other blood borne pathogens in health-care settings. MMWR; 37:377-388.

Centers for Disease Control (CDC). Guidelines for prevention of transmission of human immunodeficiency virus and hepatitis B virus to healthcare and public-safety workers. MMWR 1989; 38(S6):1-36.

Symbols

Symbols used in the labelling of this product comply with BS EN ISO 15223-1:2016 Medical Devices - 'Symbols to be used with medical device labels, labelling and information to be supplied'.

	Product Code		Single use only
	Temperature limitations		Contains sufficient for "N" tests
	Batch code		Attention, consult instructions for use
	Expiry date (last day of month)		Biohazard
	In Vitro Diagnostic device		Manufacturer
	Authorised Representative in the European Community		Test / Tests
	Components		Contents

